


From dosimeter development to routine use – Standards and Uncertainties – RAD13-38

Rolf Behrens & Oliver Hupe

[ORCID: 0000-0002-4905-7791](#)

[PTB, Department "Radiation protection dosimetry" \(6.3\)](#)

[Hyperlinks underlined and in light blue](#)



The concept of dosimetry

Standardization

- Structures
- Reference radiation fields
- Dosimeters
Type tests and Uncertainties
- Calibration and routine tests

Conclusions


The concept of dosimetry

Standardization

- Structures
- Reference radiation fields
- Dosimeters
 - Type tests and Uncertainties*
- Calibration and routine tests

Conclusions


Quantities „spread over the body“
 (finite size of organs) –
 by definition NOT measurable



ICRP


Protection quantities
 E , H_T : limits

Quantities „spread over the body“
(finite size of organs) –
by definition NOT measurable



Protection quantities
 E, H_T : limits

Point quantities (defined in
infinitesimally small point) –
by definition measurable




ICRU 4-element tissue




Measuring quantities

Quantities „spread over the body“
(finite size of organs) –
by definition NOT measurable

Point quantities (defined in
infinitesimally small point) –
by definition measurable




Protection quantities
 E, H_T : limits



Measuring quantities


Quantities „spread over the body“
(finite size of organs) –
by definition NOT measurable



ICRP

Protection quantities
 E, H_T : limits

Point quantities (defined in
infinitesimally small point) –
by definition measurable



ICRU



Measuring quantities



Devices
Indicated value

Quantities „spread over the body“
(finite size of organs) –
by definition NOT measurable

Point quantities (defined in
infinitesimally small point) –
by definition measurable



Appropriate definition of measuring (operational) quantities

Protection quantities
 E, H_T : limits

\approx

Measuring quantities

\approx

Devices
Indicated value

Type test and calibration

How can we ensure that devices measure correctly?

- **Type tests** and calibration/verification
→ **Requirements to dosimeters** in **IEC** and **ISO standards**











- **Comparable and traceable measurements/tests**
→ **Reference radiation fields** in **ISO** and **IEC standards**

The concept of dosimetry






Standardization






- **Structures** → ... *to give an overview* ...
- Reference radiation fields
- Dosimeters
Type tests and Uncertainties
- Calibration and routine tests

Conclusions

	Standards on procedures	Standards on performance requirements for instruments
International level: production of most standards	 International Organization for Standardization: <u>TC85 – SC2</u> : Radiological protection WG 2: Reference radiation fields WG 19: Individual monitoring	 International Electrotechnical Commission <u>TC45 – SC45B</u> : Radiation protection instrumentation WG 8: Active pocket and portable dose (rate) meters and monitors and passive dosimetry systems
European region: adoption of IEC and ISO standards as EN standards on a case by case decision	 European Committee for Standardization <u>CEN/TC430</u> : Nuclear energy, nuclear technologies, and radiological protection modifications of ISO standards NOT possible	 European Committee for Electrotechnical Standardization <u>CLC/TC45B</u> : Radiation protection instrumentation small modifications of IEC standards possible
National level: adoption mandatory	      ... DIN (DE), BSI (UK), AFNOR (FR), UNI (IT), UNE (ES), PKN (PL) etc.: modifications of EN standards NOT possible adoption in states of the European Union (EU), the European Free Trade Association (EFTA), Turkey (if the area is standardized in the state); contradicting national standards must be withdrawn	

	Standards on procedures	Standards on performance requirements for instruments
<p>International level: production of most standards</p>	 <p>International Organization for Standardization: <u>TC85 – SC2</u>: Radiological protection WG 2: Reference radiation fields WG 19: Individual monitoring</p>	 <p>International Electrotechnical Commission <u>TC45 – SC45B</u>: Radiation protection instrumentation WG 8: Active pocket and portable dose (rate) meters and monitors and passive dosimetry systems</p>
<p>Gulf region: adoption of IEC and ISO standards</p>	 <p>هيئة التقييس الخليجية GCC Standardization Organization https://www.gso.org.sa/en/</p>	
<p>National level: potential adoption</p>	 <p>YSMO (YE), KOWSMD (KW), QS (QA), DGSM (OM), SASO (SA), BSMD (BH), MoIAT (AE)</p>	

	Standards on procedures	Standards on performance requirements for instruments
<p>International level: production of most standards</p>	 <p>International Organization for Standardization: <u>TC85 – SC2: Radiological protection</u> WG 2: Reference radiation fields WG 19: Individual monitoring</p>	<p>International Electrotechnical Commission </p> <p><u>TC45 – SC45B:</u> <i>Radiation protection instrumentation</i> WG 8: Active pocket and portable dose (rate) meters and monitors and passive dosimetry systems</p>
<p>South Asian region: adoption of IEC and ISO standards</p>	<div style="text-align: center;">   <p>South Asian Regional Standards Organization</p> <p>https://sarso.org/</p> </div>	
<p>National level: potential adoption</p>	<div style="text-align: center;">  <p>ANSA (AF), BSTI (BD), BSB (BT), BIS (IN), MoED (MV), NBSM (NP), PSQCA (PK), SLSI (LK)</p> </div>	

	Standards on procedures	Standards on performance requirements for instruments
<p>International level: production of most standards</p>	 <p>International Organization for Standardization: <u>TC85 – SC2</u>: Radiological protection WG 2: Reference radiation fields WG 19: Individual monitoring</p>	<p>International Electrotechnical Commission </p> <p><u>TC45 – SC45B</u>: <i>Radiation protection instrumentation</i> WG 8: Active pocket and portable dose (rate) meters and monitors and passive dosimetry systems</p>
<p>African region: adoption of IEC and ISO standards</p>	  <p>https://www.arso-oran.org/</p>	<p>African Electrotechnical Standardization Commission </p> <p>https://afsec-africa.org/</p>
<p>National level: potential adoption</p>		



The concept of dosimetry



Standardization



- Structures
- **Reference radiation fields** → ***Primary and secondary standard labs (PSDL and SSDL)***
- Dosimeters
Type tests and Uncertainties
- Calibration and routine tests

Conclusions

	Photons	Neutrons	Betas
General standard	ISO 29661:2012 & Amd.1:2015: Reference radiation fields for radiation protection — Definitions and fundamental concepts — <i>Revision in progress</i>		
Characteristics and methods of production	ISO 4037-1:2019 X-rays, radionuclides, nuclear reactions	ISO 8529-1:2021 Radionuclides, nucl. react.; updated spectra	ISO 6980-1:2023 Radionuclides
Primary calibration of the fields; basic quantity	ISO 4037-2:2019 Air kerma, K_a , H	ISO 8529-2:2000 Fluence, Φ <i>Revision planned</i>	ISO 6980-2:2023 Absorbed dose, D_t Corr. factors simulated
Calibration of dosimeters and their energy and angular response; conversion coefficients from basic quantity to dose equivalent, H	ISO 4037-3:2019 $h_{pK}(0.07), h_{pK}(3), h_{pK}(10)$ $h'_{K}(0.07), h'_{K}(3), h^*_{K}(10)$	ISO 8529-3:2023 $h_{p\Phi}(10)$ $h^*_{\Phi}(10)$	ISO 6980-3:2023 $h_{pD}(0.07), h_{pD}(3)$ $h'_{D}(0.07), h'_{D}(3)$
Special considerations	ISO 4037-4:2019 low energy photons	—	—
Pulsed radiation	ISO TS 18090-1:2015 <i>Revision in progress</i>	ISO TS 18090-2 planned	—







The concept of dosimetry

Standardization

- Structures
- Reference radiation fields
- **Dosemeters** → *Manufacturers, testing labs and rad. prot. offi. / exp. (RPO/RPE)*
Type tests and Uncertainties
- Calibration and routine tests

Conclusions

Manufacturers:

- How secure (malfunction or manipulation – accidental or intentional)?
- How good must our dosimeter measure?
- What do we need to document?

Type-test laboratories:

- What to test?
- How to test?
- How to document?

Radiation protection officers / experts (RPOs/RPEs):

- What can a “type-tested” dosimeter measure?
- Does this cover my workplace (radiation type, energy, angle, temperature, ...)?
- How large is its uncertainty?

Who	What happens?	What is addressed?
Manufacturer	→ Dosemeter development (prototype)	– Characteristics and quality
Testing lab	→ Type test (a few prototype specimens)	– Relative response ∈ stated limits?
Manufacturer	→ Adjustment (each serial copy)	– Absolute response
Authority	→ Verification (each serial copy)	– Absolute response ∈ stated limits?
Exposed staff	→ Use of dosimeter	– Dose monitoring
Authority	→ Re-Verification (each serial copy)	– Absolute response ∈ stated limits?
...		

		Photons	Betas	Neutrons
Area doseimeters: $H^*(10)$, partly $H'(3)$ & $H'(0.07)$	Active	IEC 61017:2016 Environm. monitoring		IEC 61005:2014 Rate meters; revision in progress → updated techniques
		IEC 60532:2010 Fixed inst. in NPPs		
		IEC 60846-1:2009 Portable; revision in progress → $H'(3)$		IEC 61322:2020 Fixed installed
	IEC 60846-2:2015 Emergency: portable and probes			
Passive dosim. systems		IEC 62387:2020 All quantities – all types incl. hybrid doseimeters	—	
Personal doseimeters: $H_p(10)$, partly $H_p(3)$ & $H_p(0.07)$	Active	IEC 61526:2024 All types – incl. hybrid doseimeters and updated neutron requirements		
	Passive dosim. systems	IEC 62387:2020 All quantities – all types incl. hybrid doseimeters		ISO 21909-1:2021 All types ISO 21909-2:2021 Workplace considerations

List of standards available: https://www.ptb.de/cms/fileadmin/internet/fachabteilungen/abteilung_6/6.3/information/norm_lst.pdf

Software ...

- calculates dose
- indicates dose
- transmits data
- ...

... must not be changed during or after test

➔ **separate in data relevant (e.g., dose calc.)
and non-data relevant part (e.g., font, color)**

Topics addressed ...

- Identification
- Authenticity
- Data storage and transmission
- Interfaces (hard- and software)
- Documentation
- ...

		Photons	Betas	Neutrons
Area dosemeters: $H^*(10)$, partly $H'(3)$ & $H'(0.07)$	Active	IEC 61017:2016 AC; EMC; no SW		IEC 61005:2014 AC; EMC; SW
		IEC 60532:2010 AC; EMC; no SW		
		IEC 60846-1:2009 AC; EMC; SW		IEC 61322:2020 AC; EMC; no SW
		IEC 60846-2:2015 AC; EMC; SW		
Passive dosim. systems	IEC 62387:2020 AC; EMC; SW		—	
Personal dosemeters: $H_p(10)$, partly $H_p(3)$ & $H_p(0.07)$	Active	IEC 61526:2024 AC; EMC; SW		
	Passive dosim. systems	IEC 62387:2020 AC; EMC; SW		ISO 21909-1:2021 ISO 21909-2:2021 AC; no EMC ; no SW

Still standards without requirements to the software!

List of standards available: https://www.ptb.de/cms/fileadmin/internet/fachabteilungen/abteilung_6/6.3/information/norm_lst.pdf

- Standards require **mandatory (minimum) ranges**

Energy and angle, e.g., 80 keV ... 1.25 MeV → $r \in [0.71...1.67]$

Temperature, e.g., $-10\text{ °C} \dots +40\text{ °C}$ → $r \in [0.83...1.25]$

Non-linearity, e.g., 0.1 mSv ... 1 Sv → $r \in [0.87...1.18]$

...



- Relative response, r , is tested – NO absolute calibration**,
i.e., response relative to response @ reference conditions,
e.g., ^{137}Cs , 0° , 20 °C , 1 mSv ...

- Influence quantities tested one by one**
→ **presumption: independent of each other!**
→ **Additivity is required!**





Energy and angle:

- determine extreme response @ nearly mono-energ. rad.




Energy and angle:

- determine extreme response @ nearly mono-energ. rad.
- broad spectra “smear out”!




Energy and angle:


- determine extreme response @ nearly mono-energ. rad.
- broad spectra “smear out”!



- response fine for “mono-en.”?
- fine @ any spectrum & angle in documented (rated) range



Badge 1 irradiated with H_{E1} ?
 +
 Badge 2 irradiated with H_{E2} = Badge 3 irradiated with H_{E1} & H_{E2}



Badge 1 irradiated with H_{E1} ?
 +
 Badge 2 irradiated with H_{E2} = Badge 3 irradiated with H_{E1} & H_{E2}

Dosemeter construction	Method of dose calculation	Additivity fulfilled?
One detector element / signal $S \rightarrow$	dose \sim signal S	\rightarrow yes
Two or more detector elements / signals \nearrow	dose \sim linear combination or lin. optimization of signals	\rightarrow yes
	dose \sim branching algorithm, e.g., $S_1/S_2 > 1 \rightarrow$ algorithm A $S_1/S_2 \leq 1 \rightarrow$ algorithm B	\rightarrow often not \rightarrow test needed

Badge 1 irradiated with H_{E1} ?
 +
 Badge 2 irradiated with H_{E2} = Badge 3 irradiated with H_{E1} & H_{E2}

Influence quantities ... not independent of each other ...,

- e.g., **branching** (see above) → test @ mixture of radiation qualities
- e.g., **linearity depends on energy** (often the case for film dosimeters)
→ test linearity @ different energies
- e.g., **coefficient of var. depends on temperature** (can the case for active counting detectors)
→ test coefficient of variation @ different temperature
- ...


Influence quantities = Uncertainty contributions

Energy and angle: $r \in [0.71...1.67]$

Temperature: $r \in [0.83...1.25]$

Non-linearity: $r \in [0.87...1.18]$

...



IEC 62387 just fulfilled, i.e.,

Energy and angle: $r \in [0.71...1.67]$

Temperature: $r \in [0.83...1.25]$

Non-linearity: $r \in [0.87...1.18]$

...

→ $U \approx \pm 40\%$ ($k=2$; 95% cov. prob.)

IEC 62387 more than fulfilled, for example,


Energy and angle: $r \in [0.94...1.06]$

Temperature: $r \in [0.96...1.04]$

Non-linearity: $r \in [0.95...1.05]$

...

→ $U \approx \pm 10\%$ ($k=2$; 95% cov. prob.)



The concept of dosimetry


Standardization

- Structures
- Reference radiation fields
- Dosimeters
Type tests and Uncertainties
- **Calibration and routine tests**


Conclusions

ISO 14146:2024: Performance limits for individual monitoring services (IMS)

- **absolute calibration** (ph,β,n) < **factor 1.25** ≈ $R \in 0.8...1.25$: test @ reference energy
- **overall performance** (ph,β) ≲ **factor 1.5 (ICRP 75)** ≈ $R \in 0.71...1.67$: (usually) with broad spectra (routine)




- for **reference conditions** for neutrons and photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 1 mSv to 30 mSv
- for **environmental** $H^*(10)$ dosimeters for neutrons and photons (with $\bar{E} > 10$ keV): 0,035 mSv to 10 Sv
- for **workplace** $H^*(10)$ dosimeters for neutrons and photons (with $\bar{E} > 10$ keV) and for whole-body $H_p(10)$ dosimeters for photons (with $\bar{E} > 10$ keV): 0,1 mSv to 10 Sv
- for area $H'(3)$ and **eye lens** $H_p(3)$ dosimeters for photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 0,3 mSv to 10 Sv
- - - for **extremity** and whole-body $H_p(0,07)$ and area $H'(0,07)$ dosimeters for photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 1 mSv to 10 Sv and in addition for whole-body $H_p(0,07)$ dosimeters from 0,1 mSv to 1 mSv




ISO 14146:2024: Performance limits for individual monitoring services (IMS)

- **absolute calibration** (ph,β,n) < **factor 1.25** $\approx R \in 0.8...1.25$: test @ reference energy
- **overall performance** (neutrons) \lesssim **factor 2 (ICRP 75)** $\approx R \in 0.5...2.0$: (usually) with broad spectra (routine)



- for **reference conditions** for neutrons and photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 1 mSv to 30 mSv
- for **environmental** $H^*(10)$ dosimeters for neutrons and photons (with $\bar{E} > 10$ keV): 0,035 mSv to 10 Sv
- for **workplace** $H^*(10)$ dosimeters for neutrons and photons (with $\bar{E} > 10$ keV) and for whole-body $H_p(10)$ dosimeters for photons (with $\bar{E} > 10$ keV): 0,1 mSv to 10 Sv
- for area $H'(3)$ and **eye lens** $H_p(3)$ dosimeters for photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 0,3 mSv to 10 Sv
- - - for **extremity** and whole-body $H_p(0,07)$ and area $H'(0,07)$ dosimeters for photons (with $\bar{E} > 10$ keV) and betas (with $\bar{E} > 0,2$ MeV): 1 mSv to 10 Sv and in addition for whole-body $H_p(0,07)$ dosimeters from 0,1 mSv to 1 mSv




ISO 14146:2024: Performance limits for individual monitoring services (IMS)

→ **absolute calibration** (ph,β,n) < **factor 1.25** ≈ $R \in 0.8...1.25$: test @ reference energy

→ **overall performance** (ph,β) ≲ **factor 1.5** and (n) ≲ **factor 2 (ICRP 75)**

→ **demonstrated by many EURADOS (WG2,WG3) intercomparisons** (a success story...)



EURADOS →


home · events · intercomparison exercises

Intercomparison Exercises

EURADOS organizes self-sustained intercomparison exercises on a regular basis where personal dosimeters in suitable fields are irradiated and compared as well as similar exercises where detector systems for environmental monitoring are compared. EURADOS develops and disseminates documents including guidelines on a harmonized dosimetry and surveys on quality assurance. For example, recent intercomparison exercises included those

- > for **extremity** dosimeters
- > for **eye lens** dosimeters for medical applications
- > for whole body photon dosimeters on Monte Carlo-modeling for the in vivo monitoring of ^{241}Am in skull phantoms
- > for **early warning** dosimetry network systems
- > for **whole body neutron** dosimeters and
- > for **whole body photon** dosimeters.

<https://eurados.sckcen.be/en/events/intercomparison-exercises>




ISO 14146:2024: Performance limits for individual monitoring services (IMS)

→ **absolute calibration** (ph,β,n) < **factor 1.25** ≈ $R \in 0.8...1.25$: test @ reference energy

→ **overall performance** (ph,β) ≲ **factor 1.5** and (n) ≲ **factor 2 (ICRP 75)**


→ **demonstrated by many EURADOS (WG2,WG3) intercomparisons** (a success story...)



EURADOS Organisations self-sustained intercomparison exercises on a regular basis where personal dosimeters in suitable fields are irradiated and compared as well as similar exercises where detector systems for environmental monitoring are compared. EURADOS develops and disseminates documents including guidelines on a harmonized dosimetry and surveys on quality assurance. For example, recent intercomparison exercises included those

- > for **extremity** dosimeters
- > for **eye lens** dosimeters for medical applications
- > for whole body photon dosimeters on Monte Carlo-modeling for the in vivo monitoring of ^{241}Am in skull phantoms
- > for **early warning** dosimetry network systems
- > for **whole body neutron** dosimeters and
- > for **whole body photon** dosimeters.

<https://eurados.sckcen.be/en/events/intercomparison-exercises>




ISO 14146:2024: Performance limits for individual monitoring services (IMS)

→ **absolute calibration** (ph,β,n) < **factor 1.25** ≈ $R \in 0.8...1.25$: test @ reference energy

→ **overall performance** (ph,β) ≲ **factor 1.5** and (n) ≲ **factor 2 (ICRP 75)**

→ **demonstrated by many EURADOS (WG2,WG3) intercomparisons** (a success story...)



EURADOS organizes self-sustained intercomparison exercises on a regular basis where personal dosimeters in suitable fields are irradiated and compared as well as similar exercises where detector systems for environmental monitoring are compared. EURADOS develops and disseminates documents including guidelines on a harmonized dosimetry and surveys on quality assurance. For example, recent intercomparison exercises included those

- > for **extremity** dosimeters
- > for **eye lens** dosimeters for medical applications
- > for whole body photon dosimeters on Monte Carlo-modeling for the in vivo monitoring of ²⁴¹Am in skull phantoms
- > for **early warning** dosimetry network systems
- > for **whole body neutron** dosimeters and
- > for **whole body photon** dosimeters.


<https://eurados.sckcen.be/en/events/intercomparison-exercises>

Radiation Protection Dosimetry (2016), Vol. 170, No. 1–4, pp. 90–94
Advance Access publication 12 January 2016
doi:10.1093/rpd/ncv523

EURADOS PROGRAMME OF INTERCOMPARISONS FOR INDIVIDUAL MONITORING SERVICES: SEVEN YEARS OF DEVELOPMENT AND FUTURE PLANS

T. Grimbergen^{1,*}, M. Figel², A. McWhan³, A. M. Romero⁴ and H. Stadmann⁵
¹NRG, Individual Monitoring, PO Box 9034, 6800 ES Arnhem, The Netherlands
²Helmholtz Zentrum München, Auswertungsstelle, Otto-Hahn-Ring 6, D-81739 München, Germany
³Cavendish Nuclear Limited, Berkeley Approved Dosimetry Service, A11 Berkeley Centre, Berkeley, 4, Gloucestershire GL139FB, UK
⁴CIEMAT, Radiation Dosimetry, Avda Complutense 40, 28040 Madrid, Spain
⁵Seibersdorf Labor GmbH, Radiation Protection Dosimetry, A-2444 Seibersdorf, Austria

*Corresponding author: grimbergen@nrg.eu



Question	Yes (%)	No (%)	Not appl. (%)
Did the results from the intercomparison help you to improve your service (to your own judgement)?	~85	~10	~5
Did the results from the intercomparison help you to pass inspections or audits?	~80	~15	~5
Do you intend to participate in other intercomparisons in the future?	~100	0	0

<https://doi.org/10.1093/rpd/ncv523>

The concept of dosimetry

Standardization

- Structures
- Reference radiation fields
- Dosemeters
Type tests and Uncertainties
- Calibration and routine tests

Conclusions

- Production of most standards at international level
- Adoption at regional and national level → (e.g., EN and DIN) standards
→ contribution at international level is most influential
- Type-test standards → demonstrate performance of a dosimeter in rated ranges
→ RPO/RPE: workplace in rated ranges?
- Calibration @ reference energy or workplace spectrum advisable
 - Uncertainty ($k=2$; 95 % cov. prob.) \in ICRP 75 ($R_{\text{photon,beta}} \lesssim$ factor 1.5; $R_{\text{neutron}} \lesssim$ factor 2))!
- Overall performance to be demonstrated (intercomparisons)
- List of standards is available at PTB's website: http://www.ptb.de/cms/fileadmin/internet/fachabteilungen/abteilung_6/6.3/information/norm_lst.pdf

From dosimeter development to routine use – Standards and Uncertainties – RAD13-38

Rolf Behrens & Oliver Hupe

[ORCID: 0000-0002-4905-7791](https://orcid.org/0000-0002-4905-7791)

[PTB, Department "Radiation protection dosimetry" \(6.3\)](#)

[Hyperlinks underlined and in light blue](#)



**Physikalisch-Technische Bundesanstalt
Braunschweig and Berlin**

Bundesallee 100
38116 Braunschweig



GERMANY

Dr. Rolf Behrens
Phone: +49 531 592-6340
E-mail: Rolf.Behrens@PTB.de

www.ptb.de


06/2025




From dosimeter development to routine use – Standards and Uncertainties – RAP25-16

Rolf Behrens & Oliver Hupe

Additional material
Quantities in radiation protection



ICRP 116 – 2010




E
Effective dose


$$= \sum_T W_T \cdot \sum_R W_R$$

Tissue- & Radiation-weighting factors


$D_{T,R}$
Absorbed dose in organs



For the calculation of conversion coefficients



Definition in reference phantoms



H
Dose equivalent


$$= Q(L)$$

Quality factor in tissue

D
Absorbed dose in tissue





Definition in sphere or person:
Area- and personal dose



H
Dose equivalent


= $Q(L)$
Quality factor
in tissue

· D
Absorbed dose
in tissue




$$Q_{\text{photon};\text{beta}} = 1$$

$$1 < Q_{\text{neutron}} < 30$$



- Q : Quality factor** to take into account the biological effectiveness depending on the quality of the radiation
- $Q(L)$ is a function of a physical quantity
 - L is the linear energy transfer (in keV/μm) in water
 - L can be measured with Tissue Equivalent Proportional Counters (TEPC)



$$H = Q(L) \cdot D$$


Dose equivalent = Quality factor in tissue · Absorbed dose in tissue




$$\text{For representation} = h$$


Conversion coefficient

$$\cdot \Phi \text{ bzw. } K_a$$




Fluence or air kerma in air




	Whole body	Lens of the eye	Local skin
Protection quantities (ICRP 116)	 <p>ICRP reference voxel phantoms: $E_{\text{eff}} = \sum_T w_T \sum_R w_R D_{T,R}$</p>	 <p>Stylized eye model; whole lens (ICRP 116, Annex F): $H_{\text{lens}} = \sum_R w_R D_{\text{lens},R}$</p>	 <p>Tissue-equivalent cube (10x10x10 cm³); 1 cm² area at 50 – 100 μm depth (ICRP 116, Annex G): $H_{\text{local skin}} = \sum_R w_R D_{\text{local skin},R}$</p>












Sievert



	Whole body	Lens of the eye	Local skin
Protection quantities (ICRP 116) 	ICRP reference voxel phantoms: $E_{\text{eff}} = \sum_T w_T \sum_R w_R D_{T,R}$	 Stylized eye model; whole lens (ICRP 116, Annex F): $H_{\text{lens}} = \sum_R w_R D_{\text{lens},R}$	 Tissue-equivalent cube (10x10x10 cm ³); 1 cm ² area at 50 – 100 μm depth (ICRP 116, Annex G): $H_{\text{local skin}} = \sum_R w_R D_{\text{local skin},R}$



Operational quantities: definition: $H = Q(L) \cdot D$

Sievert


	Area	Area	Area
Operational quantities for monitoring (ICRU 51/57)	 ICRU 4-element tissue sphere: $\varnothing = 30$ cm: $H^*(10) = Q \cdot D(10)_{\text{sph}}$	 ICRU 4-element tissue sphere: $\varnothing = 30$ cm: $H'(3;\Omega) = Q \cdot D(3;\Omega)_{\text{sph}}$	 ICRU 4-element tissue sphere: $\varnothing = 30$ cm: $H'(0.07;\Omega) = Q \cdot D(0.07;\Omega)_{\text{sph}}$
	Individual  $H_p(10) = Q \cdot D(10)_{\text{person}}$ 	 $H_p(3) = Q \cdot D(3)_{\text{person}}$ 	 $H_p(0.07) = Q \cdot D(0.07)_{\text{pers.}}$ 
	For calibration: ICRU 4-elem. tiss. slab: 30x30x15 cm ³ : $H_p(10) = Q \cdot D(10)_{\text{slab}}$	For calibration: ICRU 4-elem. t. cylinder: $\varnothing = h = 20$ cm: $H_p(3) = Q \cdot D(3)_{\text{cylinder}}$	For calibration: ICRU 4-el. tissue slab, pillar, rod ($\varnothing = 73, 19$ mm): $H_p(0.07) = Q \cdot D(0.07)_{\text{slab, pillar, rod}}$




From dosimeter development to routine use – Standards and Uncertainties – RAP25-16

Rolf Behrens & Oliver Hupe

Additional material
Phantoms at a glance ...





Phantoms for the calculation of protection quantities

- Anthropomorphic voxel, eye, and skin phantoms from ICRP
- To calculate the absorbed dose to the organ and the effective dose

Phantoms for the calculation of operational quantities

- Defined by the ICRU, consisting of ICRU 4-element tissue
- To calculate the conversion coefficients
- ICRU sphere (30 cm diameter) for area dosimetry
- ICRU slab/cylinder/pillar/rod phantom for personal dosimetry
- No realization required



(For calculation of conversion coefficients for calibrations)

Phantoms for type tests and calibrations

- defined by ISO, made from PMMA and water
- Simulate the backscattered radiation field
- ISO water slab phantom (with PMMA walls)
- ISO water cylinder phantom (with PMMA walls)
- ISO water pillar phantom (with PMMA walls)
- ISO PMMA rod phantom
- Calibration of area dosimeters *without* phantoms

